

| | | | |
|------------------------------|-----------------------------|------------------------|---|
| X12B SAFETY CHECKLIST | | | BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE |
| Rev: C | Effective: 5/24/2010 | Page 1 of 1 | Number: LS-SCL-0026 |
| Prepared by: Philip Marino | | Reviewed by: P. Marino | Approved by: B. Chmiel |

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

UPSTREAM OF X12A HUTCH

- 1. Valid Padlock Index and log, no padlocks open.
- 2. Bremsstrahlung shield (BS #1) in place and banded.
- 3. Bremsstrahlung shield (BS #2) in place and banded.
- 4. Slit tank wrapped in lead. Leaded glass on Observation Port 1.
- 5. Water cooled Beryllium window #1 wrapped in lead.
- 6. Bremsstrahlung shield 3 (BS #3) intact.
- 7. Leaded glass on Observation Port 2 (OP2).
- 8. Exclusion Zone (EZ #1) running between OP2 and BS4 intact and secure.
- 9. Bremsstrahlung shield (BS #4) in place and banded.

IN X12A HUTCH

- 10. Exclusion Zone (EZ #2), wall to wall black foam, in place and intact.

BETWEEN X12A AND X12C HUTCH

- 11. Exclusion Zone (EZ #3) before Monochromator intact; Bellows 3 under EZ3 wrapped in lead.
- 12. Monochromator tank wrapped in lead; Mono Observation ports 3 & 4 covered with lead glass.
- 13. White beam stop assembly wrapped in lead.
- 14. Vacuum Bellows #4 on multilayer tank wrapped in lead.
- 15. Exclusion Zone (EZ # 4) and Bremsstrahlung shield (BS # 5) in place and intact.
- 16. Vacuum Bellows wrapped in painted lead.
- 17. Mirror tank viewport covered with lead glass.
- 18. Vacuum Bellows wrapped in painted lead.

IN X12C HUTCH

- 19. Exclusion Zone (EZ #4) wall to wall intact (In-line bellows 7 wrapped in lead).

BETWEEN X12C AND X12B HUTCH

- 20. Exclusion Zone (EZ #5) between X12C hutch and Photon shutter intact.
- 21. Beam Observation ports 6&7 covered with lead glass.
- 22. Photon shutter lead shielding intact.
- 23. Bremsstrahlung shield (BS #6) in place and covered.

IN X12B HUTCH

- 24. Bremsstrahlung shield (BS #7) on downstream hutch wall intact and covered.

OPCO / Beamline Staff.: _____ Date: _____